

*Roche*

October 15, 2020

To Whom It May Concern:


Subject: Self Declaration

Blood Glucose Meter Accu-Chek® Guide Me Supplied by Roche Diabetes Care.

Herewith we certify that the above meter for the purpose of monitoring glucose concentration in human blood are tested according to Environmental Conditions and Test Procedures for Airborne Equipment (RTCA DO-160G) Section 21 Emission of Radio Frequency Energy.

When the meter is in flight mode, the meter fulfills the requirements according to the standard, category M.

The meter does not affect the performance of aircrafts.

A handwritten signature in blue ink, which appears to read "Gregory Smith", followed by the date "10/15/2020". The signature is stylized and somewhat cursive.

Gregory Smith

Vice President of Quality and Regulatory  
Roche Diabetes Care, Inc.