



March 21, 2014

To Whom It May Concern:

Subject: SELF DECLARATION

Blood Glucose Meter ACCU-CHEK® Aviva Connect Supplied by Roche Diagnostics.

Herewith we certify that the above meter for the purpose of monitoring glucose concentration in human blood are tested according to Environmental Conditions and Test Procedures for Airborne Equipment (RTCA/DO-160E) Section 21 Emission of Radio Frequency Energy.

When the meter is in flight mode, the meter fulfills the requirements according to this standard, category M.

The meter does not affect the performance of aircrafts.

A handwritten signature in black ink, appearing to read "Tony Heleine".

Tony Heleine
Principal R&D Engineer
Diabetes Care Systems

3/21/2014

A handwritten signature in black ink, appearing to read "John Krug".

John Krug
Director of Quality
Diabetes Care Systems

21 March 2014