



August 9, 2012

To Whom It May Concern:

Subject: Self Declaration

**Blood glucose meters: ACCU-CHEK® Aviva and ACCU-CHEK® Nano  
supplied by Roche**

Herewith I certify that the above meters for the purpose of monitoring glucose concentration in human blood are tested according to Environmental Conditions and Test Procedures for Airborne Equipment (RTCA/DO-160) Section 21 Emission of Radio Frequency Energy.

The meters fulfill the requirements according to this standard, category M.

The meters do not affect the performance of aircraft.

A handwritten signature in black ink, appearing to read "Karen Roberts".

Karen Roberts  
Director, Design Quality  
Roche Diabetes Care

A handwritten signature in black ink, appearing to read "Ann Stankiewicz".

Ann Stankiewicz  
Interim Director, Quality  
Roche Diabetes Care

CADMS 33116

**Roche Diagnostics Corporation**  
9115 Hague Road  
Indianapolis, IN 46256  
+1-317-521-2000  
USA