



March 21, 2014

To Whom It May Concern:

**Subject: SELF DECLARATION**

**Blood Glucose Meter ACCU-CHEK® Aviva Expert Supplied by Roche Diagnostics.**

Herewith we certify that the above meter for the purpose of monitoring glucose concentration in human blood are tested according to Environmental Conditions and Test Procedures for Airborne Equipment (RTCA/DO-160E) Section 21 Emission of Radio Frequency Energy.

The meter fulfills the requirements according to this standard, category M.

The meter does not affect the performance of aircrafts.

A handwritten signature in black ink, appearing to read "Tony Heleine".

Tony Heleine  
Principal R&D Engineer  
Diabetes Care Systems

3/21/2014

A handwritten signature in black ink, appearing to read "John Krug".

John Krug  
Director of Quality  
Diabetes Care Systems

21 March 2014